



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/976,639	10/11/2001	Takehiko Shoji	56232.8 [4829]	2040
7590	11/06/2003		EXAMINER	
Squire, Sanders & Dempsey L.L.P. Suite 300 One Maritime Plaza San Francisco, CA 94111			MORAN, TIMOTHY J	
			ART UNIT	PAPER NUMBER
			2878	

DATE MAILED: 11/06/2003

Determination of Patent Term Adjustment under 35 U.S.C. 154 (b)

(application filed on or after May 29, 2000)

The Patent Term Adjustment to date is 261 day(s). If the issue fee is paid on the date that is three months after t mailing date of this notice and the patent issues on the Tuesday before the date that is 28 weeks (six and a ha months) after the mailing date of this notice, the Patent Term Adjustment will be 261 day(s).

If a Continued Prosecution Application (CPA) was filed in the above-identified application, the filing date th determines Patent Term Adjustment is the filing date of the most recent CPA.

Applicant will be able to obtain more detailed information by accessing the Patent Application Information Retrieval (PAIR) system (<http://pair.uspto.gov>).

Any questions regarding the Patent Term Extension or Adjustment determination should be directed to the Office Patent Legal Administration at (703) 305-1383. Questions relating to issue and publication fee payments should directed to the Customer Service Center of the Office of Patent Publication at (703) 305-8283.

Notice of Allowability	Application No.	Applicant(s)
	09/976,639	SHOJI ET AL.
	Examiner	Art Unit
	Timothy J. Moran	2878

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTO-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. This communication is responsive to *amendment filed October 16, 2003*.
2. The allowed claim(s) is/are 7,9,11-15 and 17-28.
3. The drawings filed on 11 October 2001 are accepted by the Examiner.
4. Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) All Some* None of the:
 1. Certified copies of the priority documents have been received.
 2. Certified copies of the priority documents have been received in Application No. _____.
 3. Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

5. Acknowledgment is made of a claim for domestic priority under 35 U.S.C. § 119(e) (to a provisional application).
 - (a) The translation of the foreign language provisional application has been received.
6. Acknowledgment is made of a claim for domestic priority under 35 U.S.C. §§ 120 and/or 121.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application. **THIS THREE-MONTH PERIOD IS NOT EXTENDABLE**

7. A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.

8. CORRECTED DRAWINGS must be submitted.
 - (a) including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) hereto or 2) to Paper No. _____.
 - (b) including changes required by the proposed drawing correction filed _____, which has been approved by the Examiner.
 - (c) including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No. _____.

Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet.

9. DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1 <input type="checkbox"/> Notice of References Cited (PTO-892)	2 <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)
3 <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)	4 <input checked="" type="checkbox"/> Interview Summary (PTO-413), Paper No. <u>enc1</u> .
5 <input type="checkbox"/> Information Disclosure Statements (PTO-1449), Paper No. _____.	6 <input type="checkbox"/> Examiner's Amendment/Comment
7 <input type="checkbox"/> Examiner's Comment Regarding Requirement for Deposit of Biological Material	8 <input type="checkbox"/> Examiner's Statement of Reasons for Allowance
	9 <input type="checkbox"/> Other

EXAMINER'S AMENDMENT

An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Charles Runyan on October 29, 2003.

The application has been amended as follows:

In claim 15, the term "comprises a thermo-welding resins on surface" is replaced with --comprises a thermo-welding resin on its surface--.

In claim 24, the term "of claim 19" is replaced with --of claim 23--.

This amendment is necessary to ensure clarity of the claims.

Allowable Subject Matter

Claims 7, 9, 11-15, and 17-28 are allowed.

The following is an examiner's statement of reasons for allowance:

Claim 7 includes the limitations that a protective film in a stimulable phosphor panel has a water vapor transmission rate of not more than 10 g/m² per day. Claim 9 includes the limitations that a protective film in a stimulable phosphor panel comprises two resin layers and a stimulating light absorbing layer between them. Claim 11 includes limitations that the surface roughness of a protective film in a stimulable phosphor panel is controlled as stated in the claim.